

APR 27 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130660 ✓
Name	DARYL RICE	ID#	2088 Date 04/19/09
A	Agency	SITKA POLICE DEPT.	Phone # 747-3245
Instrument Location		304 LAKE STREET	
Alco S/N		X124617 ✓	Target Value .085 High Pressure 900
B			
Alco Test Values		.091 ✓	.091 ✓
		1 st Alco	2 nd Alco
Signature		Daryl Rice	003 5/13/09
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 12 day of May, 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130660 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Daryl Rice
04/19/09
WASP
900

ALASKA DEPARTMENT OF PUBLIC SAFETY
DIVISION OF CRIME PREVENTION

APRIL 19, 2009

OPERATOR'S NAME: *Daryl Rice*
NAME DARYL RICE
OPERATOR'S NUMBER: 900

VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

SEX: M
BIRTHDATE: 01/01/71
BIRTHPLACE: YUC
TEST TYPE: V
TARGET VALUE: .085
ALCO 1/10 X124617

--- BREATH ANALYSIS ---

AGE ADJUSTED FOR 30.25 in

ALCO TARGET	.086	22:06
BLANK TEST	.000	22:07
INTERNAL STANDARD	VERIFIED	22:07
ALCO IV 30.25 in	.091	22:07
BLANK TEST	.000	22:08
SUBJECT SAMPLE	.000	22:08
BLANK TEST	.000	22:09
ALCO IV 30.25 in	.086	22:09
BLANK TEST	.000	22:10

DIAGNOSTIC CHECK

Daryl Rice
04/19/09
WASP

ALASKA DEPARTMENT OF PUBLIC SAFETY
DIVISION OF CRIME PREVENTION

APRIL 19, 2009 ✓
TIME 22:03

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

NUMBERS

SAMPLE CHAMBER: 50c

WARM TIME: 41c

DETECTOR: 200X in

FLAME DETECTOR: OKAY

FLAME SPEED: OKAY

DETECTOR: OKAY

FILMS: OKAY

INTERNAL STANDARD: OKAY

VERIFICATION: OKAY

PRINTER TEST

ALCO 1/10 X124617